



# अखिल भारतीय आयुर्वेद संस्थान ALL INDIA INSTITUTE OF AYURVEDA (AIIA)

(आयुष मंत्रालय, भारत सरकार के अंतर्गत स्वायत्त संस्थान)  
(An Autonomous Organization under the Ministry of AYUSH, Govt. of India)

F.No. AIIA/AB/20005/1/1/2016-AIIA/RRDR/VOL.III/123

Dated 04/02/2021

## ADDENDUM/CORRIGENDUM

Tender Ref No. AIIA/AB/20005/1/1/2016-AIIA/RRDR/VOL.III/UV-VIS Spectrophoto Meter

Dated: 14/01/2021

Title: Tender Documents for Supply Installation, Testing and Commissioning of UV-VIS Spectrophoto Meter for Department of RRDR, AIIA-Delhi

Below Mentioned Amendments has been issued in Above mentioned NIT and Tender Documents.

1. As per instruction available page no. 2 and Page no. 5 (Para 1 (j) ) of NIT/ Tender Document the goods should be domestic or already imported by the bidder on their own arrangement, any provisions/conditions/clause contained in this tender document in contradiction to the aforesaid provisions should be treated as withdrawn.
2. Tender document fees has been withdrawn
3. EMD/ Bid Security for this tender has been withdrawn. However, Bidders to sign a Bid securing declaration accepting that if they withdraw or modify their Bids during the period of validity, or if they are awarded the contract and they fail to sign the contract, or to submit a performance security before the deadline defined in the request for bids document, they will be suspended for the Two Years from being eligible to submit Bids for contracts with All India Institute of Ayurveda New Delhi.

S No	Page No Clause No	Current Clause	Revised Clause.
1	Page No 25 Clause No 3	Detector: Silicon photodiode with double detector separately for sample and reference respectively.	Specifications of Detector may be changed to "Photomultiplier or silicon photodiode"
2	Page No 25 Clause No 8	Spectral Bandwidth: Upto 1.4 nm or better.	Spectral Bandwidth (point no.8) may be corrected to "0.1-5 nm or better".

Therefore, all the concerned are here by informed that tender no. **AIIA/AB/20005/1/1/2016-AIIA/RRDR/VOL.III/ UV-VIS Spectrophoto Meter** having ID. 2021\_AIIA\_582538\_1 floated on 14/01/2021 for the procurement of **UV-VIS Spectrophoto Meter** Due to amendments in terms and condition/Specifications mentioned in tender document.

Hence date of submission & opening bid is now 15/02/2021 at 1500 & 1530 Respectively this is in place of date and time mentioned in NIT/Tender document dated 14/01/2021.

Consequently, above mentioned corrigendum/addendums become the integral part of the tender document, rest of the constants remains same.

  
(Joint Director)

गौतमपुरी, मथुरा रोड, नई दिल्ली-110044

Gautampuri, Mathura Road, New Delhi-110044

ई-मेल/E-Mail : director@aiia.gov.in दूरभाष/Phone : 011-26950401/402