

अखिल भारतीय आयुर्वेद संस्थान

ALL INDIA INSTITUTE OF AYURVEDA (AIIA)

(आयुष मंत्राखय, आश्त शरकार के अंतर्गत स्वायत्त शंस्थान) (An Autonomous Organization under the Ministry of AYUSH, Govt. of India)

F.No. 64/Store/Shalya Instrument/AIIA-2020/2914

Dated - 17 02 2021

ADDENDUM/CORRIGENDUM

Tender Ref No 64/Store/Shalya Instrument/AllA-2020/ Video Gastro-Colonoscope Complete set / General surgery items dated: 10/02/2021

<u>Title: Tender Documents for Supply Installation, Testing and Commissioning of Video Gastro-Colonoscope Complete</u>
<u>set / General surgery items for Department of Shalya Tantra, AliA-Delhi</u>

The below Mentioned Amendments has been issued in Above mentioned NIT and Tender Documents.

Colonosco re	Jltra-high esolution HDTV	Now the following is added. Ultra-high resolution HDTV imaging capability for clearer
complete set cl ol of firm arm set uring light coordinate coordina	assage even at ktreme angles. pecification for blonoscope:	observation of fine capillaries and subtle mucosal structures. Uniform image brightness-even under difficult lighting conditions. Optimal instrument passage even at extreme angles 1: Video Processor: Should be compatible with Analog, HD-SDI and DVI output & 16:9 & 16:10 Output for a HDTV monitor should be available. 2: LED light source: should contain the long-life. - Equipped with high resolution HDTV Imaging capacity. - Compact, lightweight (10-11 kg) and ergonomically designed - Narrow Band Imaging capacity for compatibility with NBI Video scopes. - Equipped with one touch connection of scopes. - Portable Memory & USB Slot for image recording - Automatic IRIS control & automatic white balance - Equipped with memory back up for settings & Lithium battery. - Should have pre freeze function for image stabilization 3. Medical grade Trolley suitable for Endoscope System -1no 4. Monitor- Medical Grade its should not be less then 22inch or more-1no 5. Reporting Software 1no 6. Latest i5, 16GB ram 01 TB Hard Disc with latest Intel processor with 9th generation this system should not has base frequency less then 1.8 gHZ.Computer System and Printer for reporting- 1SET

Page 1 of 2

<u>किंशागिक</u>

शौतमपुरी, मथुरा शेड, नई दिल्ली-110044

Gautampuri, Mathura Road, New Delhi-110044

ई-मेल/E-Mail: director@aiia.gov.in दूरभाष/Phone: 011-26950401/402

The following items should be provided with Video Gastro Colonoscope

- Carrying Case
- ETO Cap
- Leakage Tester
- Caps for Working
- Channel
- Irrigation Tube
- Instrument Oil, 50 ml
- Bite protector (in gastroscope only)
- Biopsy

Forceps, oval cup

- Cleaning Brush
- Cleaning Valve
- Video Connecting Cable

The following items should be provided with Video Gastro Colonoscope

- Carrying Case
- ETO Cap
- Leakage Tester
- Caps for Working Channel
- Irrigation Tube
- Instrument Oil, 50 ml
- Bite protector (in gastroscope only)
- Biopsy Forceps, oval cup
- Cleaning Brush
- Cleaning Valve
- Video Connecting Cable

White cap holder-1 No.

Foot Holder-1 No.

Scope cable holder-1 No.

SDI cable 2.5M-1 No.

Keyboard-1 No.

Portable memory (2GB)-1 No.

Keyboard cover-1 No.

Water Container-1 No.

Operation Manual-1 No.

White balance cap-1 No.

Note: Practical training to consultants. It should be US FDA/ CE/BIS certified.

Therefore, all the concerned are here by informed that tender no_64/Store/Shalya Instrument/AIIA-2020/ Video Gastro-Colonoscope Complete set having ID. 2021_AIIA_588847_1 floated on 10/02/2021 for the procurement of Video Gastro-Colonoscope Complete Set Due to amendments in Specifications mentioned in tender document.

Consequently, above mentioned corrigendum/addendums become the integral part of the tender document, rest of the constants remains same.

(Joint Director)