



अखिल भारतीय आयुर्वेद संस्थान
ALL INDIA INSTITUTE OF AYURVEDA (AIIA)
(आयुष मंत्रालय, भारत सरकार के अंतर्गत स्वायत्त संस्थान)
(An Autonomous Organization under the Ministry of AYUSH, Govt. of India)

F.No. AIIA/B/20005/1/1/2016-AIIA/RRDR/Vol.III/Fourier Transform Infrared Spectroscopy (FTIR)

Dated: 10/02/2021

ADDENDUM/CORRIGENDUM

Tender Ref No: AIIA/B/20005/1/1/2016-AIIA/RRDR/Vol.III/Fourier Transform Infrared Spectroscopy (FTIR) dated: 10/02/2021


Title: Tender Documents for Supply, Installation, Testing and Commissioning of Fourier Transform Infrared Spectroscopy (FTIR)

The below Mentioned Amendments has been issued in Above mentioned NIT and Tender Documents.

S No	Page No Clause No	Current Clause	Revised Clause.
1	Page No. 20, point No. 5	Signal-to-Noise ratio: 48,000:1 for 1-minute scan at 4 cm-1 or better	Signal-to-Noise ratio: 30,000:1 for 1-minute scan at 4 cm-1 or better
2	Page No. 20, Point No. 9	Optics: All mirrors must be gold coated or equivalent coating for high throughput and to give sensitivity of 48,000:1 for 1-minute scan peak to peak	Optics: All mirrors must be gold coated or equivalent coating for high throughput and to give sensitivity of 30,000:1 for 1-minute scan peak to peak
3	Page No. 20, Point No. 11	Vendor has to provide the latest model of (Del /HP) make computer and laser jet colour printer available on the date of bidding and submit their technical details in the bid	Vendor has to provide the latest model of Del /HP make computer, UPS, and laser jet colour printer available on the date of bidding and submit their technical details in the bid

Therefore, all the concerned are here by informed that tender no **AIIA/B/20005/1/1/2016-AIIA/RRDR/Vol.III/Fourier Transform Infrared Spectroscopy (FTIR)** having ID. **2021_AIIA_589031_1** floated on **10/02/2021** for the procurement of **Fourier Transform Infrared Spectroscopy (FTIR)** has been amended.

Consequently, above mentioned corrigendum/addendums become the integral part of the tender document, rest of the constants remains same.


(Joint Director)



Email :director-aiia@gov.in

Page 1 of 1
गौतमपुरी, सरिता विहार, मथुरा रोड, नई दिल्ली-110076 -
Gautampuri, Sarita Vihar, Mathura road, New Delhi -110076
Phone :011-29948658

Fax :011-299948660