



अखिल भारतीय आयुर्वेद संस्थान

ALL INDIA INSTITUTE OF AYURVEDA (AIIA)

(आयुष मंत्रालय, भारत सरकार के अंतर्गत स्वायत्त संस्थान)

(An Autonomous Organization under the Ministry of AYUSH, Govt. of India)

F.No. 64/Store Instrument/AIIA-2020/2915

Dated - 17/02/2021

ADDENDUM/CORRIGENDUM

Tender Ref No 64/Store Instrument/AIIA-2020/ Laparoscope Complete set + SOS items dated: 09/02/2021

Title: Tender Documents for Supply Installation, Testing and Commissioning of Laparoscope Complete Set + SOS Items for Department of Shalya Tantra, AIIA-Delhi

The below Mentioned Amendments has been issued in Above mentioned NIT and Tender Documents.

Laparoscope Complete set + Hand instrument set + SOS items Mentioned on page No. 21 of tender document	LAPROSCOPY COMPLETE SET: 01 SET	LAPROSCOPY COMPLETE SET: 01 SET
	1. Video camera - Video camera- Full HD Vision with CMOS chips Technology with in-built full HD recording of Videos and Still Images. Auto exposure zoom seakale coupler model	1. (i) Video camera Video camera- Full HD video image including light source with 3 Chip CCD/ CMOS Full HD Camera Head. The camera head should be compatible for ICG. (ii) A separate medical grade recorder with full HD recording of Images & videos with recording capability of 500 GB or more. It should have USB option. FTP facility, full HD1920 *1080 PIXEL
	2. Display Unit: Display Unit - LED Monitor (19inch)	2. Display Unit: Medical Grade LEDMonitor26 inch or more.
	3. Electronic hi flow Insufflator 45-50Litter/min	3. Electronic hi flow Insufflator 45-50Litter/min Digital display for flow rate, pneumo-peritoneum pressure and volume of gas consumed
	4. LED light source with equivalent Lux of 300W Xenon.= (02)	4. LED light source with equivalent Lux of 300W Xenon.= (01)
	6. F.O. cable 4.8 mm 230 cm Long	5. F.O. cable 4.8 mm 230 cm Long-02 Practical training to consultants. Video Image processor, Camera Head, Light source, Light Guide Cable and Telescope should be US FDA/ CE/BIS certified and should be from same manufacturer for system compatibility.
S.No. 21	Vessel sealing 04	Deleted

Therefore, all the concerned are here by informed that tender no **64/Store Instrument/AIIA-2020/ Laparoscope Complete set + SOS items** having ID. **2021_AIIA_588815_1** floated on 09/02/2021 for the procurement of **Laparoscope Complete set + SOS items** Due to amendments in terms and condition/Specifications mentioned in tender document.

Consequently, above mentioned corrigendum/addendums become the integral part of the tender document, rest of the constants remains same.


(Joint Director)