



अखिल भारतीय आयुर्वेद संस्थान ALL INDIA INSTITUTE OF AYURVEDA (AIIA)

(आयुष मंत्रालय, भारत सरकार के अंतर्गत स्वायत्त संस्थान)
(An Autonomous Organization under the Ministry of AYUSH, Govt. of India)

F.No. 66/Store/SRPT Instruments/AIIA-2021

12929

Dated: 10.02.2021

ADDENDUM/CORRIGENDUM

Tender Ref No: 66/Store/SRPT Instruments/AIIA-2021_Normal Birthing Simulator

Dated: 10/02/2021

Title: Tender Documents for Supply Installation, Testing and Commissioning of Normal Birthing Simulator

The below Mentioned Amendments has been issued in Above mentioned NIT and Tender Documents.


S No	Page No Clause No	Current Clause	Revised Clause.
1	Tender Document page No 20	<ol style="list-style-type: none">1. The simulator that is ideal for demonstrating the mechanics of various birthing positions and train skills including abdominal, vaginal examinations, normal birth, vacuum assisted delivery, shoulder dystocia and breech.2. It can be used as a table top model for demonstration and skills training and it is ideal for in-situ team simulation and training respectful care.3. Skills trainer with anatomically correct landmarks.4. The simulator offers very unique features, yet it is significantly more affordable than other simulators.5. It is made of durable fabrics, requires no lubrication during use, and has passed durability test for 3000 deliveries.6. Simulator should compromise the following: Placenta w/o	<ol style="list-style-type: none">1. The simulator should be ideal for demonstrating the mechanics of various birthing positions and train skills including abdominal, vaginal examinations, normal birth, vacuum assisted delivery, shoulder dystocia and breech.2. It can be used as a table top model for demonstration and skills training and it should be ideal for in-situ team simulation and training respectful care.3. Skills trainer with anatomically correct landmarks.4. It should be made of durable fabrics, requiring no lubrication during use, and should have passed the durability test for deliveries.5. Birthing Simulator should have the facilities to train the following: Fetal Heart rate monitoring,

ओ.प.न.स.

		umbilical cord and membranes, 2 Ties for umbilical cord, 2 pairs of gloves , Fertoscope, Urine catheter , 20 ml syringe, Fetal skull with fontanelles,Cervix Cups (4, 6 & 8 cm dilations , BabyBirthie Newborn Simulator (dark or light complexion), Baby Birthie hat, Two sheets to simulate towels, Table clamp ,Carrying bag for Mama Birthie	Vaginal delivery, Breech delivery, Vacuum delivery, Incomplete Placenta. 6. Simulator should compromise the following: Placenta with umbilical cord and membranes, 2 Ties for umbilical cord, 2 pairs of gloves , Fetoscope, Urine catheter , 20 ml syringe, Fetal skull with fontanelles,Cervix Cups (4, 6 & 8 cm dilations , Baby Newborn Simulator (dark or light complexion), Baby hat, Two sheets to simulate towels, Table clamp ,Carrying bag for the simulator
--	--	--	--

Therefore, all the concerned are here by informed that tender no 66/Store /SRPT Instruments/AIIA-2021/ Normal Birthing Simulator having ID. 2021_AIIA_589053_1 floated on 10/02/2021 for the procurement of Normal Birthing Simulator. Due amendments in terms and condition/Specifications mentioned in tender document have been ammended.

Consequently, above mentioned corrigendum/addendums become the integral part of the tender document, rest of the constants remains same.


 (Joint Director)