



अखिल भारतीय आयुर्वेद संस्थान
ALL INDIA INSTITUTE OF AYURVEDA (AIIA)
(आयुष मंत्रालय, भारत सरकार के अंतर्गत स्वायत्त संस्थान)
(An Autonomous Organization under the Ministry of AYUSH, Govt. of India)

F.No. G-19004/5/1/2016-AIIA / 2920 -

Dated- 17/2/2021

ADDENDUM/CORRIGENDUM

Tender Ref No: G-19004/5/1/2016-AIIA/ Deep Freezers for Molecular

Dated: 10/02/2021

Title: Tender Documents for Supply Installation, Testing and Commissioning of Deep Freezers for Molecular biology & Cell culture Laboratory, AIIA-DELHI

The below Mentioned Amendments has been issued in Above mentioned NIT and Tender Documents.

S No	Page No Clause No	Current Clause	Revised Clause.
1.	Tender Document page No 20	Deep Freezer -20 degree celcius <ul style="list-style-type: none">Capacity (l): more than 230It should have 4 to 6 shelves of stainless steel with 4 to 5 inner doors."Should have defrost: auto defrostSettable range should be: -19 to -21 C	Deep Freezer -20 degree celcius <ul style="list-style-type: none">Capacity (l): more than 300 LIt should have 4 to 6 shelvesShould have defrost: auto/manual defrostSettable range should be: -19/-20 to -21/-30°C
2	Tender Document page No 20 & 21	Deep Freezer -80 degree celcius <ul style="list-style-type: none">Freezer should hold 600 2" boxes or 60,000 2ml Vials and should be min of 800-900 liters capacityThe freezer should be constructed using 1" thick vacuum panel insulation in conjunction with environmentally-friendly water blown foam.	Deep Freezer -80 degree celcius <ul style="list-style-type: none">Freezer should hold 550 or more 2" boxes or 50000 or more 2ml Vials and should be min of 800-900 liters capacityThe freezer should be constructed using vacuum insulation panel in conjunction with environmentally friendly PUF insulation or water blown foam.

Page 1 of 2

गौतमपुरी, सरिता विहार, मथुरा रोड, नई दिल्ली -110076

Gautampuri, Sarita Vihar, Mathura road, NEW DELHI-110076

E-mail: aiiانewdelhi@gmail.com

Phone: 011-29948658

Fax: 011-29948660



अखिल भारतीय आयुर्वेद संस्थान

ALL INDIA INSTITUTE OF AYURVEDA (AIIA)

(आयुष मंत्रालय, भारत सरकार के अंतर्गत स्वायत्त संस्थान)

(An Autonomous Organization under the Ministry of AYUSH, Govt. of India)

	<ul style="list-style-type: none">Freezer should have at least 4-6 internal storage compartments with a minimum of 4-6 polystyrene insulated inner doors to ensure sample security. Inner doors have no latches or external magnets and must be removable for easy cleaning without the use of tools.Natural refrigerants should compliant with EU F-Gas regulation: R170 and R290It should be FDA Class II Medical Device	<ul style="list-style-type: none">Freezer should have at least 4-6 internal storage compartments with a 2 or more polystyrene insulated inner doors to ensure sample security. Inner doors may/ may not have latches or external magnets and must be removable for easy cleaning with/without the use of tools.Freezer should use only natural, commercially available refrigerants (Hydrocarbon/CFC/HCFC Free) refrigerants with no special blends required.It should be CE/FDA Class II Medical Device
--	--	--

Therefore, all the concerned are here by informed that tender no - G-19004/5/1/2016-AIIA/ Deep Freezers for Molecular biology & Cell culture Laboratory having tender ID 2021_AIIA_589091_1 floated on 10/02/2021 for the procurement of microscopes is amended as above. Consequently, above mentioned corrigendum/addendums become the integral part of the tender document, rest of the constants remains same.

(Joint Director)