



अखिल भारतीय आयुर्वेद संस्थान ALL INDIA INSTITUTE OF AYURVEDA (AIIA)

(आयुष मंत्रालय, भारत सरकार के अंतर्गत स्वायत्त संस्थान)
(An Autonomous Organization under the Ministry of AYUSH, Govt. of India)

Our Ref: F.20005/2/3/2020-AIIA/COVID-1 | 3462

Date. 20/03/2021

Limited Tender Enquiry No- F.20005/2/3/2020-AIIA/COVID-19

Dear M/s

A limited Tender Enquiry is hereby floated invited for purchasing the following items required by the Department of RNVVAIIA, New Delhi.

S.No	Items Description (Specification and Description. annexed)	Quantity required (Pack of 96 Wells)
1.	IL-6	4
2.	IL-10	4
3.	IL-2	4
4.	IL-12	4
5.	IL-4	4
6.	Gamma Interferon	4
7.	IgG	4
8.	IgM	4
9.	IgA	4
10.	Vitamin-D	3

Technical Specifications Annexed - Annexure- I

Terms and Conditions

The quotations should be addressed to **The Director, All India Institute of Ayurveda, New Delhi and sent by post / Courier by 26/03/2021 up to 15:00 hrs. Quotations should be sealed in an envelope (i.e. LTE No F.20005/2/3/2020-AIIA/COVID- 19 for the purchase of ELISA KITS).**

1. The firm must be registered and having TIN No / GST registration no (please mention TIN / GST Registration no. on the bids) or else the Quotations will be rejected.

The Bidder is required to submit bid in two parts viz

A. Technical bid in one envelope containing following information.

- i. Technical description of Quoted Model/product.
- ii. Original catalogue of the Quoted Product
- iii. All required Affidavits / undertakings
- iv. Principal company Authorization Letter

गौतमपुरी, मथुरा रोड, नई दिल्ली-110044

Gautampuri, Mathura Road, New Delhi-110044

ई-मेल/E-Mail : director@aia.gov.in दूरभाष/Phone : 011-26950401/402

- B. Financial bid in one envelope containing following information
- i. Offered price in INR only of quoted model / Kits
 - ii. Price list of all consumables / accessories required to run the equipment.

BOTH BIDS SHOULD PACAED IN SEPARATE SEALED ENVELOPS

2. EMD- Bidders need not to deposit EMD/ Bid Security for this tender. However, Bidders to sign a Bid securing declaration accepting that if they withdraw or modify their Bids during the period of validity, or if they are awarded the contract and they fail to sign the contract, or to submit a performance security before the deadline defined in the request for bids document, they will be suspended for the Two Years from being eligible to submit Bids for contracts with All India Institute of Ayurveda New Delhi.
3. The offered make/brand should be mentioned on the bid. Relevant literature /original catalogue should be submitted along with the technical bid filing which item would not be considered.
4. Quotation should be typed only. No overwriting or erased entries should be there in the quotation. The rates should be valid for at least nine months.
5. GST or any other kind of tax(s) must be separately mentioned against each item. In case no sales tax is chargeable, price must be quoted as NET PRICE.
6. Demonstration: The purchaser reserves the right to ask for a free demonstration of the quoted equipment/product at a pre-determined place acceptable to the purchaser for technical acceptability as per the tender specification, before the opening of the Price Bid.
7. Warranty /Guaranty: The item must be covered under 1 year comprehensive warranty (Including all spares and labor) form the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected.
8. Shelf life of the product to be supplied is left at least 2/ 3 of the total shelf life at the time of supply

Undertakings:

- An undertaking should be enclosed with the technical bid state that “the firm has deposited the GST to the GST department of last period and there is no outstanding of taxes is pending against the firm.”
- The firm has to give an affidavit on non-judicial stamp paper that the firm has not been black listed in the past by any Institution, Government/ Private.
- The firm has to give and affidavit on non-judicial stamp paper that there is no vigilance FEMA/CBI case pending against the firm/ supplier
- The firm has to give an affidavit on a non-judicial stamp paper that the firm is not supplying the same item at lower rates quoted in the tender to any government organization or any other institution during past 1 year, as per FALL CLAUSE adhered by DGS&D and other government agencies
- The firm has to give an affidavit that “We hereby certify that if at any time, information furnished by us is proved to be false or incorrect, we are liable for any action as deemed fit by the purchaser.

Installation and training of Kits /equipments will be provided by the company at free of cost and there will no extra charges be paid.

The firm must be able to supply the item at the mentioned place/ location within stipulated time for supply after issuing the Supply Order

Payment will be made through ECS/RTGS/NEFT after delivery and inspection of working conditions of the item. Please provide following information on the bill

1. Name of the beneficiary
2. Account no. of the beneficiary
3. Bank/Branch name
4. IFSC code of the bank/branch
5. An undertaking on each page of the bill state what "It is certified that the firm is depositing GST to GST department quarterly and filling the returns in stipulated time and there are no dues of GST against the firm" without undertaking the bill will not be processed for payment.

The institute is not authorized to issue C/D form against the order placed in Indian Rupees.

The successful bidder firm has to be deposit a 10% performance Bank Guarantee of total value of items, in favour of The Director, All India Institute of Ayurveda, New Delhi" in form of Account pay DD/FDR/Bank Guarantee should be valid for two additional months on warranty periods, before releasing the payment.

It has been advised to all bidder firms that they may please check the Institute's web-site www.aiia.gov.in time to time if any corrigendum issuing regarding this L.T.E.

The quantity of items can be increased or decreased in the extent of 25%

The Director, All India Institute of Ayurveda reserves the rights to cancel/reject full or any part of the rate enquiry which generally do not fulfill the conditions stipulated in the rate enquiry.



Director

ALL INDIA INSTITUTE OF AYURVEDA (AIIA)

Gautam Puri, Sarita Vihar

New Delhi 110076

Web Site: www.aiia.gov.in

21/11/28

SPECIFICATIONS OF ELISA KITS – 96 WELLS

Specification of total IgG Elisa kit

1. Kit should be CE certified and IVD approved.
2. Kit must give quantitative result in biological fluids as serum or plasma.
3. There should be at least 4 to 5 calibrators and one control inside the kit.
4. Sensitivity: minimum detection limit of the kit should be ≤ 0.15 g/l.
5. Specificity: specificity of the kit should not be less than 99%.
6. Linearity: 0-20 g/l or more

Specification of Total IgM Elisa kit

1. Kit should be CE certified and IVD approved.
2. Kit must give quantitative results in biological fluids serum, plasma, urine etc.
3. There should be at least 4 to 5 calibrators and one control inside at kit.
4. Sensitivity: minimum detection limit of the kit should be ≤ 0.15 g/l.
5. Specificity: specificity of the kit should not be less than 99%.
6. Linearity 0-10 g/l or more

Specification of total IgA Elisa kit

1. Kit should be CE certified and IVD approved.
2. Kit must give quantitative result in biological fluids as serum or plasma.
3. There should be at least 4 to 5 calibrators and one control inside the kit.
4. Sensitivity: minimum detection limit of the kit should be ≤ 0.15 g/l.
5. Specificity: specificity of the kit should not be less than 99%.
6. Linearity: 0-5 g/l or more

Specification of Interlukin-2 (IL-2) Elisa Kit

1. Kit should be IVD approved 96 well ELISA kit
2. Kit must give quantitative result in serum or plasma sample.
3. There should be at least 4 to 5 calibrators and 1 control inside the kit.
4. Sensitivity minimum detection limit of the kit should ≤ 7.0 pg/ml.
5. Specificity - no cross reaction with other interleukins.
6. Range: till 1000 pg/ml or more

Specification of IL-4 Elisa kit

1. Kit should be IVD approved 96 well ELISA kit
2. Kit must give quantitative result in serum or plasma sample.
3. There should be at least 4 to 5 calibrators and 1 control inside the kit.
6. Sensitivity of the kit should be ≤ 0.5 pg/ml.
7. Specificity : no cross reaction with other interleukins

Richa



SPECIFICATIONS OF ELISA KITS – 96 WELLS

8. Linearity till 100 pg/ml or more

Specification of IL-6 Elisa kit

1. Kit should be IVD approved 96 well ELISA kit
2. Kit must give quantitative result in serum or plasma sample.
3. There should be at least 4 to 5 calibrators and 1 control inside the kit.
4. Sensitivity minimum detection limit of kit should be ≤ 1.0 pg/ml.
5. Specificity no cross reactions with other cytokines.
6. Linearity upto 250pg/ml or more

Specification of gamma-interferon Elisa kit

1. Kit should be IVD approved 96 well ELISA kit
2. Kit must give quantitative result in serum or plasma sample.
3. There should be at least 4 to 5 calibrators and 1 control inside the kit.
4. Sensitivity minimum detection limit of the kit should be ≤ 2.0 pg/ml.
5. Specificity no cross reaction with other cytokines
6. Linearity till 900 pg/ml or more

Specification of IL-10 Elisa kit

1. Kit should be IVD approved 96 well ELISA kit
2. Kit must give quantitative result in serum or plasma sample.
3. There should be at least 4 to 5 calibrators and 1 control inside the kit.
4. Sensitivity minimum detection limit of the kit should be ≤ 2 pg/ml.
5. Specificity no cross reactivity with other cytokines.
6. Linearity till 450pg/ml or more

Specification of IL-12 Elisa kit

1. Kit should be IVD approved 96 well ELISA kit
2. Kit must give quantitative result in serum or plasma sample.
3. There should be at least 4 to 5 calibrators and 1 control inside the kit.
4. Sensitivity minimum detection limit of the kit should be ≤ 10 pg/ml.
5. Specificity no cross reactivity with other cytokines.
6. Linearity till 900pg/ml or more

Richa

[Signature]