



अखिल भारतीय आयुर्वेद संस्थान ALL INDIA INSTITUTE OF AYURVEDA (AIIA)

(आयुष मंत्रालय, भारत सरकार के अंतर्गत स्वायत्त संस्थान)

(An Autonomous Organization under the Ministry of AYUSH, Govt. of India)

F. No. M-15/21/2021-AIIA

Date: November 2021

ADDENDUM/CORRIGENDUM

Tender Ref No. M-15/21/2021-AIIA

Title: Tender Documents for Supply, Installation, Testing and Commissioning of Laser Doppler Blood Flow Meter

With reference to the tender of even no. having **Tender ID 2021_AIIA_628597_1** floated on 13th October 2021 for the procurement of Laser Doppler Blood Flow Meter, it is hereby informed to all concerned that following amendments have been issued in respect to the above-mentioned NIT and Tender document:

Sl. No.	Page No. Clause No.	Current Clause	Revised Clause
1.	Tender document page no -21-23	The system should have an inbuilt TFT/LED display for instant monitoring of the oximetry data	The system should have an inbuilt TFT/LED display or direct PC connectivity for the instant monitoring of the oximetry data
2.	Tender document page no -21-23	Accessories provides along with Device >Standard Probe >Digit Probe >Miniature Surface Probe (sutureable and non-sutureable) >Standard Needle Probe >Fine Needle Probe >Bent Tip Needle Probe >Adhesive Rings for surface probes >Miniature Holders for needle probes >Calibration Kit for INL191 probes	Accessories provides along with Device *Surface probes *Standard, Needle probes with miniature holders *Adhesive rings for surface probes *Calibration kit for all probes Other probes may be procured in future with same price as quoted.

Further, it is also to inform that bid submission end date has been extended for eight days i.e. 11th November 2021.

Consequently, above mentioned corrigendum/addendums become the integral part of the tender document, rest of the contents remains same.

(Dr. Umesh Tagade)
Joint Director

गौतमपुरी, सरिता विहार, मथुरा रोड, नई दिल्ली -110076
Gautampuri, Sarita Vihar, Mathura road, NEW DELHI-110076

E-mail: director-aiia@gov.in

Phone: 011-29948658

Fax: 011-29948660