



अखिल भारतीय आयुर्वेद संस्थान

ALL INDIA INSTITUTE OF AYURVEDA (AIIA)

(आयुष मंत्रालय, भारत सरकार के अंतर्गत स्वायत्त संस्थान)

(An Autonomous Organization under the Ministry of AYUSH, Govt. of India)

F. No. O-50/2/2021-AIIA-Part(3)

Date: December 2021

ADDENDUM/CORRIGENDUM

Tender Ref No. O-50/2/2021-AIIA-Part(3)

Title: Tender Documents for Gas Chromatography-Tantam Mass Spectrometer

With reference to the tender of even no. having **Tender ID 2021_AIIA_634635_1** floated on 25th November 2021 for the procurement of **Gas Chromatography-Tantam Mass Spectrometer**, it is hereby informed to all concerned that following amendments have been issued in respect to the above-mentioned NIT and Tender document:

Sl. No	Page No. Clause No.	Current Clause	Revised Clause
1.	Tender document Page no. 19 Point no. 5	<p>Ion Source</p> <p>It should have EI source. Programmable to 350 °C or more</p> <p>It should have Integrated, dual filament for EI & CI with improved filament lifetime</p> <p>Dedicated Ion source for EI and CI and removable with requires a breaking in the vacuum and the source can be cleaned after removing from instrument with venting the vacuum</p>	<p>Ion Source</p> <p>It should have EI source. Programmable to 350 °C or more</p> <p>It should have Integrated, dual filament for EI & CI with improved filament lifetime</p> <p>Dedicated Ion source for EI and CI</p>
2.	Tender document Page no. 20 point no. 18	<p>Injector/Inlets</p> <p>Split/Splitless with electronics gas controller Qty 1</p> <p>PTV injector- Qty 1 for large/Full volume injection</p> <p>The system should have column end or mid column backflush to remove unwanted components/ contaminants/ high boilers.</p>	<p>Injector/Inlets</p> <p>Split/Splitless with electronics gas controller Qty 1</p> <p>PTV/MMI injector- Qty 1 for large/Full volume injection</p> <p>The system should have column end or mid column backflush to remove unwanted components/ contaminants/ high boilers</p>

Further, Bid Submission Date and Bid Opening Date have also been extended for seven days in the above-mentioned NIT and Tender as per the following details:

Contd. on page 2

गौतमपुरी, सरिता विहार, मथुरा रोड, नई दिल्ली -110076

Gautampuri, Sarita Vihar, Mathura road, NEW DELHI-110076

E-mail: director-aiia@gov.in

Phone: 011-29948658

Fax: 011-29948660

I/1230/2021

- 2 -

Bid Submission Date:

For: 16-Dec-2021

Read: 23-Dec-2021

Bid Opening Date:

For: 16-Dec-2021

Read: 23-Dec-2021

Consequently, above mentioned corrigendum/addendum become the integral part of the tender document, rest of the contents remain same.

(Dr. Umesh Tagade)
Joint Director